

FINAL REPORT - ACUTE HEMOLYSIS

Test Article:

Elastomer Shell

Conclusion:

The Test Article, assayed as described above, was found to be negative for Hemolytic Activity; Test Article passes the Acute Hemolysis Test.

FINAL REPORT - ACUTE HEMOLYSIS

Test Article: Gel

Conclusion:

The Test Article, assayed as described above, was found to be negative for Hemolytic Activity; Test Article passes the Acute Hemolysis Test.

FINAL REPORT - ACUTE HEMOLYSIS

Test Article:

Leaf Valve Assembly:

Conclusion:

The Test Article, assayed as described above, was found to be ~~negative~~ for Hemolytic Activity; Test Article ~~passes~~ the Acute Hemolysis Test.